

WD/WC Certificate

<p>Customer Information</p> <p>Name: Customer Name Here PO: 123456789</p>	<p>Manufacturer Information</p> <p>Sales Order: 1234567</p>
--	---

Model Number	Rosemount Serial Number	Tag Number	UNIQUE ID NUMBER
1595P030A3LA065WCQC1Q8	0030043	3EPT-FRC-606B	9996516037

1595P040A3LA065WCQC1Q8	0030043	3EPT-FRC-603B	9996516217
------------------------	---------	---------------	------------

This certifies that Rosemount, Inc. parts, on the above referenced Sales Order and Customer Purchase Order, conform to all applicable specifications for option WC and/or WD Flow Calibration Verification.

WC = Discharge Coefficient Verification (3 point)
WD = Discharge Coefficient Verification (full 10 point)

The conditioning orifice plate models 1595 and 405C are designed by Dieterich Standard, Inc. (an Emerson Process Management/Rosemount company), an ISO 9001 ANSI/ASQC Q9001 accredited company. We hereby certify that the product(s) supplied conform to approved policies, procedures, and specifications for the manufacture of the equipment.

Lab flow calibration verification report(s) are attached.



Tim Layer
Vice-President of Global Quality

WC-WD Flow Calibration Verification attachment here