



TECHNICAL STANDARDS &  
SAFETY AUTHORITY  
4th Floor, West Tower  
3300 Bloor Street West  
Toronto, Ontario  
Canada M8X 2X4



## STATUTORY DECLARATION Registration of Fittings

I, Michael Zinn, Approval Engineer  
(Name and Position, e.g. President, Plant Manager, Chief Engineer)

of Micro Motion, Inc.  
(Name of Manufacturer)

Located at 7070 Winchester Circle, Boulder, CO 80301 USA 303-530-8554  
(Plant Address) (Telephone No.) (Fax No.)

do solemnly declare that the fittings listed hereunder, which are subject to the **Technical Standards and Safety Act**, Boilers and Pressure Vessels Regulation, comply with all of the requirements of ASME B31.3

(Title of recognized North American Standard)

which specifies the dimensions, materials of construction, pressure/temperature ratings, identification marking the fittings and service;

or are not covered by the provisions of a recognized North American standard and are therefore manufactured to comply with Micro Motion Published Data as supported by the attached data which identifies the dimensions, material of construction, pressure/temperature ratings and the basis for such ratings, the marking of the fitting for identification and service.

I further declare that the manufacture of these fittings is controlled by a quality system meeting the requirements of ISO 9001 which has been verified by the following authority, DNV

The items covered by this declaration, for which I seek registration, are category F type fittings. In support of this application, the following information and/or test data are attached as follows:

F025 Tube Drawing, Tube Spec, Calculations, Product Data Sheet


(drawings, calculations, test reports, etc.)

Declared before me at Boulder County in the State of Colorado  
the 8th day of June AD 2004.

Commissioner for Oaths:

Christina Kae Foust  
(Printed name)

Christina Kae Foust  
(Signature)



Michael Zinn  
(Signature of Declarer)


To the best of my knowledge and belief, the application meets the requirements of the **Technical Standards and Safety Act**, Boilers and Pressure Vessels Regulation, and CSA Standard B51 and is accepted for registration in my Commission Expires 05/08/2008

CRN: OF9473-5

Registered by: K. Russell

Dated: JUNE 25/04

NOTE: This registration expires on JUNE 25 / 2014



NOTE: 1 HAWP = 217 PSIG / Full Vac @ 257  
K.R