

Mass Flowmeter Calibration Certificate

Micro Motion, Inc.
7070 Winchester Circle
Boulder, CO 80301



Serial Number: 490529
Equipment Model: CMF100
Calibration Stand: TSM3A
Location: Boulder, CO
Test File Name / No.: 1.20568461
Date of Calibration: 7/30/2006

Procedure Used: CP 79


Received Status:

<input checked="" type="checkbox"/>	In Tolerance, no adjustment required (Error < 0.035%)
<input type="checkbox"/>	In Tolerance, adjustment required (0.035% < Error < 0.055%)
<input type="checkbox"/>	Out of Tolerance* (0.055% < Error)

*Requires immediate notification of Director of Quality

Returned Status:

<input checked="" type="checkbox"/>	In Tolerance
<input type="checkbox"/>	Removed from Use

Calibration Performed By: Steven Seek
SSC Calibration Engineer
Signature: 
GRM S/N: 467065

Reviewed By: Mark Lee
Manager, Reliability Engineering
Signature: 

Mass Flowmeter Calibration Certificate

Micro Motion, Inc.
7070 Winchester Circle
Boulder, CO 80301



Serial Number: 490637
Equipment Model: CMF200
Calibration Stand: TSM3A
Location: Boulder, CO
Test File Name / No.: 1.20568457
Date of Calibration: 7/30/2006

Procedure Used: CP 79

Received Status:

<input checked="" type="checkbox"/>	In Tolerance, no adjustment required (Error < 0.035%)
<input type="checkbox"/>	In Tolerance, adjustment required (0.035% < Error < 0.055%)
<input type="checkbox"/>	Out of Tolerance* (0.055% < Error)

*Requires immediate notification of Director of Quality

Returned Status:

<input checked="" type="checkbox"/>	In Tolerance
<input type="checkbox"/>	Removed from Use

Calibration Performed By: Steven Seek
SSC Calibration Engineer
Signature: 
GRM S/N: 11005153

Reviewed By: Mark Lee
Manager, Reliability Engineering
Signature: 

Mass Flowmeter Calibration Certificate

Micro Motion, Inc.
7070 Winchester Circle
Boulder, CO 80301



Serial Number: 11005158
Equipment Model: CMF300
Calibration Stand: TSM3A
Location: Boulder, CO
Test File Name / No.: 1.20568464
Date of Calibration: 7/30/2006

Procedure Used: CP 79

Received Status:

<input checked="" type="checkbox"/>	In Tolerance, no adjustment required (Error < 0.035%)
<input type="checkbox"/>	In Tolerance, adjustment required (0.035% < Error < 0.055%)
<input type="checkbox"/>	Out of Tolerance* (0.055% < Error)

*Requires immediate notification of Director of Quality

Returned Status:

<input checked="" type="checkbox"/>	In Tolerance
<input type="checkbox"/>	Removed from Use

Calibration Performed By: Steven Seek
SSC Calibration Engineer
Signature: 
GRM S/N: 11021217

Reviewed By: Mark Lee
Manager, Reliability Engineering
Signature: 

Mass Flowmeter Calibration Certificate

Micro Motion, Inc.
7070 Winchester Circle
Boulder, CO 80301



Serial Number: 411024
Equipment Model: CMF-400 (A)
Calibration Stand: TSM3A
Location: Boulder, CO
Test File Name / No.: 1.205690014
Date of Calibration: 7/31/2006

Procedure Used: CP 79

Received Status:


<input checked="" type="checkbox"/>	In Tolerance, no adjustment required (Error < 0.035%)
<input type="checkbox"/>	In Tolerance, adjustment required (0.035% < Error < 0.055%)
<input type="checkbox"/>	Out of Tolerance* (0.055% < Error)

*Requires immediate notification of Director of Quality

Returned Status:

<input checked="" type="checkbox"/>	In Tolerance
<input type="checkbox"/>	Removed from Use

Calibration Performed By: Steven Seek
SSC Calibration Engineer
Signature: 
GRM S/N: 11021217

Reviewed By: Mark Lee
Manager, Reliability Engineering
Signature: 

Mass Flowmeter Calibration Certificate

Micro Motion, Inc.
7070 Winchester Circle
Boulder, CO 80301



Serial Number: 411564
Equipment Model: CMF400 (B)
Calibration Stand: TSM3A
Location: Boulder, CO
Test File Name / No.: 1.20569153
Date of Calibration: 7/31/2006

Procedure Used: CP 79

Received Status:

<input checked="" type="checkbox"/>	In Tolerance, no adjustment required (Error < 0.035%)
<input type="checkbox"/>	In Tolerance, adjustment required (0.035% < Error < 0.055%)
<input type="checkbox"/>	Out of Tolerance* (0.055% < Error)

*Requires immediate notification of Director of Quality

Returned Status:

<input checked="" type="checkbox"/>	In Tolerance
<input type="checkbox"/>	Removed from Use

Calibration Performed By: Steven Seek
SSC Calibration Engineer
Signature: 
GRM S/N: 11021217

Reviewed By: Mark Lee
Manager, Reliability Engineering
Signature: 

Mass Flowmeter Calibration Certificate

Micro Motion, Inc.
7070 Winchester Circle
Boulder, CO 80301



Serial Number: 411565
Equipment Model: CMF400 (C)
Calibration Stand: TSM3A
Location: Boulder, CO
Test File Name / No.: 1_20569216
Date of Calibration: 7/31/2006

Procedure Used: CP 79

Received Status:

<input checked="" type="checkbox"/>	In Tolerance, no adjustment required (Error < 0.035%)
<input type="checkbox"/>	In Tolerance, adjustment required (0.035% < Error < 0.055%)
<input type="checkbox"/>	Out of Tolerance* (0.055% < Error)

*Requires immediate notification of Director of Quality

Returned Status:

<input checked="" type="checkbox"/>	In Tolerance
<input type="checkbox"/>	Removed from Use

Calibration Performed By: Steven Seek
SSC Calibration Engineer
Signature:

Reviewed By: Mark Lee
Manager, Reliability Engineering
Signature:

Standards Used: GRM S/N: 11021217