

This is to certify that the following meter was calibrated in accordance with procedures **LWI-112** and **LWI-115** for performing primary gravimetric calibrations in the Micro Motion Measurement Technology Test Lab. The meter has been calibrated with standards whose accuracies are traceable to N.I.S.T.

Micro Motion, Inc. accredited by the American Association for Laboratory Accreditation (A2LA) for the specific scope of accreditation under Certificate Number 2033.01.

**STANDARDS USED**

19F5-19F9, 19FA-19FN, 19FP-19FZ, 19GC-19GK, 19GL, 74876, 586735, 590294, 590295, 51-2006939

**ESTIMATED UNCERTAINTY  $\pm 0.014\%$**

The uncertainty is estimated using a coverage factor (k) of 2, providing a confidence level of approximately 95 %

**STATUS**

As Found

As Left

  
**TEST TECHNICIAN**

12/07/2006  
**DATE**

  
**QUALITY ASSURANCE REVIEW**

12/7/06  
**DATE**

**TEST DATE & TIME**

11/30/2006 4:57:28 PM

**TEST COMMENTS**

CMF300 Global Reference Meter (Initial Calibration) 315pt Flow Test 7K

**SENSOR DATA**

Serial Number: 14022514  
Model: ELITE  
Size: 300  
Material: 316L Stainless Steel  
Press Rating: STANDARD  
Temp Rating: STANDARD

**SENSOR CALIBRATION**

Flow Calibration Factor: 698.984.45  
D1: 0.00000  
D2: 1.00000  
K1: 10519.93  
K2: 12479.67  
FD: 210.10  
Density Temp Coeff.: 4.45

**TRANSMITTER DATA**

Serial Number: 09001145  
Model: MVD - 2400S  
Mass Flow Units: lb/min  
Density Units: g/cc  
Temperature Units: C  
Mass Flow Cutoff: 5.00  
Mass Flow Damping: 0.6  
Flow Direction: FWD  
Frequency Flow Rate: 7000  
Frequency Span: 10000

**METER SPECIFICATIONS**

Flow Accuracy: 0.10 %  
Flow Repeatability: 0.05 %  
Zero Stability: 0.25000 lb/min  
Density Accuracy: 0.0005 g/cc  
Density Repeatability: 0.0002 g/cc

